

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086093.D
 Acq On : 25 Mar 2025 12:15
 Operator : JC\MD
 Sample : Q1618-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP20250320-02

Quant Time: Mar 25 15:25:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	350362	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131617	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139878	54.508	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.020%
35) Dibromofluoromethane	8.165	113	124167	50.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.565	98	426665	48.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.847	95	155820	49.229	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086093.D
Acq On : 25 Mar 2025 12:15
Operator : JC\MD
Sample : Q1618-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP20250320-02

Quant Time: Mar 25 15:25:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

