

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086096.D
 Acq On : 25 Mar 2025 13:28
 Operator : JC\MD
 Sample : PB167276ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#01

Quant Time: Mar 26 03:16:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318766	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126469	55.219	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.440%		
35) Dibromofluoromethane	8.165	113	113240	50.896	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.800%		
50) Toluene-d8	10.565	98	412778	51.118	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.240%		
62) 4-Bromofluorobenzene	12.847	95	120455	41.828	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 83.660%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	33222	47.480	ug/l	96
20) Methylene Chloride	5.283	84	13464	6.578	ug/l	96
43) Isopropyl Acetate	8.688	43	118196	22.069	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086096.D
Acq On : 25 Mar 2025 13:28
Operator : JC\MD
Sample : PB167276ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#01

Quant Time: Mar 26 03:16:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

