

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086097.D
 Acq On : 25 Mar 2025 13:52
 Operator : JC\MD
 Sample : PB167276ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#02

Quant Time: Mar 26 03:17:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	331060	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	130297	54.704	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.400%		
35) Dibromofluoromethane	8.165	113	116070	50.230	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.460%		
50) Toluene-d8	10.565	98	428386	51.081	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.160%		
62) 4-Bromofluorobenzene	12.847	95	133356	44.588	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 89.180%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	31064	42.689	ug/l	99
20) Methylene Chloride	5.277	84	13194	6.198	ug/l	96
43) Isopropyl Acetate	8.688	43	124701	22.418	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086097.D
Acq On : 25 Mar 2025 13:52
Operator : JC\MD
Sample : PB167276ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#02

Quant Time: Mar 26 03:17:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

