

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086098.D
 Acq On : 25 Mar 2025 14:16
 Operator : JC\MD
 Sample : PB167276ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#03

Quant Time: Mar 26 03:17:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	351402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138215	53.939	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.880%		
35) Dibromofluoromethane	8.165	113	120766	49.237	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.480%		
50) Toluene-d8	10.565	98	451844	50.759	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.520%		
62) 4-Bromofluorobenzene	12.847	95	135364	42.639	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 85.280%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	30456	38.904	ug/l	100
20) Methylene Chloride	5.283	84	13099	5.720	ug/l	98
43) Isopropyl Acetate	8.688	43	132121	22.377	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086098.D
Acq On : 25 Mar 2025 14:16
Operator : JC\MD
Sample : PB167276ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#03

Quant Time: Mar 26 03:17:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

