

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086102.D
 Acq On : 25 Mar 2025 15:52
 Operator : JC\MD
 Sample : PB167276ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#07

Quant Time: Mar 26 03:19:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	317156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134096	55.642	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.280%	
35) Dibromofluoromethane	8.159	113	117795	50.701	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.400%	
50) Toluene-d8	10.565	98	423135	50.182	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.360%	
62) 4-Bromofluorobenzene	12.847	95	127947	42.548	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	85.100%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	85533	116.170	ug/l	100
20) Methylene Chloride	5.271	84	6771	3.144	ug/l	98
25) 2-Butanone	7.483	43	11684	10.998	ug/l	90
43) Isopropyl Acetate	8.689	43	55414	9.908	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086102.D
Acq On : 25 Mar 2025 15:52
Operator : JC\MD
Sample : PB167276ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#07

Quant Time: Mar 26 03:19:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

