

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086103.D
 Acq On : 25 Mar 2025 16:16
 Operator : JC\MD
 Sample : PB167276ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#08

Quant Time: Mar 26 03:20:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	328374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	131417	55.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.580%
35) Dibromofluoromethane	8.165	113	113715	49.614	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.220%
50) Toluene-d8	10.565	98	414613	49.843	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.847	95	124814	42.073	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.140%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	85184	118.371	ug/l	98
20) Methylene Chloride	5.271	84	6502	3.088	ug/l #	96
25) 2-Butanone	7.488	43	11305	10.887	ug/l	89
43) Isopropyl Acetate	8.688	43	50835	9.214	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086103.D
Acq On : 25 Mar 2025 16:16
Operator : JC\MD
Sample : PB167276ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#08

Quant Time: Mar 26 03:20:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

