

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086105.D
 Acq On : 25 Mar 2025 17:04
 Operator : JC\MD
 Sample : PB167276ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167276ZHE#10

Quant Time: Mar 26 03:21:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109871	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124797	55.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.100%
35) Dibromofluoromethane	8.165	113	109006	49.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	10.565	98	399803	49.607	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.847	95	119932	41.727	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.460%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	87770	126.738	ug/l	100
20) Methylene Chloride	5.289	84	6008	2.966	ug/l #	89
25) 2-Butanone	7.483	43	10796	10.804	ug/l	93
43) Isopropyl Acetate	8.683	43	53198	9.952	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086105.D
Acq On : 25 Mar 2025 17:04
Operator : JC\MD
Sample : PB167276ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167276ZHE#10

Quant Time: Mar 26 03:21:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

