

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086108.D
 Acq On : 25 Mar 2025 18:16
 Operator : JC\MD
 Sample : Q1619-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Mar 26 03:23:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	127963	55.457	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.920%		
35) Dibromofluoromethane	8.165	113	112378	50.486	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.980%		
50) Toluene-d8	10.565	98	412211	51.025	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.060%		
62) 4-Bromofluorobenzene	12.847	95	127289	44.181	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.360%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	27243	38.645	ug/l	97
20) Methylene Chloride	5.271	84	10702	5.189	ug/l	91
43) Isopropyl Acetate	8.682	43	100568	18.769	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086108.D
Acq On : 25 Mar 2025 18:16
Operator : JC\MD
Sample : Q1619-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Mar 26 03:23:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

