

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086110.D
 Acq On : 25 Mar 2025 19:04
 Operator : JC\MD
 Sample : Q1619-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Mar 26 03:25:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318327	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	311916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126987	56.908	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.820%
35) Dibromofluoromethane	8.165	113	110682	49.815	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	10.565	98	414122	51.355	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.720%
62) 4-Bromofluorobenzene	12.847	95	134691	46.836	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.680%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	28679	42.068	ug/l #	86
20) Methylene Chloride	5.277	84	11291	5.662	ug/l	93
43) Isopropyl Acetate	8.682	43	108038	20.200	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086110.D
Acq On : 25 Mar 2025 19:04
Operator : JC\MD
Sample : Q1619-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Mar 26 03:25:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

