

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086111.D
 Acq On : 25 Mar 2025 19:28
 Operator : JC\MD
 Sample : Q1626-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CO-32-1

Quant Time: Mar 26 03:25:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	313069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129226	56.517	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.040%	
35) Dibromofluoromethane	8.165	113	111924	51.220	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.440%	
50) Toluene-d8	10.565	98	405905	51.182	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.360%	
62) 4-Bromofluorobenzene	12.847	95	130113	46.004	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	76626	109.693	ug/l	97
20) Methylene Chloride	5.265	84	4898	2.397	ug/l	92
25) 2-Butanone	7.483	43	11198	11.110	ug/l #	78
43) Isopropyl Acetate	8.688	43	33709	6.408	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086111.D
Acq On : 25 Mar 2025 19:28
Operator : JC\MD
Sample : Q1626-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CO-32-1

Quant Time: Mar 26 03:25:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

