

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086113.D
 Acq On : 25 Mar 2025 20:16
 Operator : JC\MD
 Sample : Q1630-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-206

Quant Time: Mar 26 03:26:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	320819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	309527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	132112	58.051	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.100%	
35) Dibromofluoromethane	8.165	113	112767	50.359	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.720%	
50) Toluene-d8	10.565	98	412728	50.785	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.560%	
62) 4-Bromofluorobenzene	12.847	95	133528	46.071	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	68996	99.235	ug/l	95
20) Methylene Chloride	5.271	84	5591	2.749	ug/l #	91
25) 2-Butanone	7.477	43	9539	9.508	ug/l #	81
43) Isopropyl Acetate	8.683	43	26691	4.952	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086113.D
Acq On : 25 Mar 2025 20:16
Operator : JC\MD
Sample : Q1630-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-206

Quant Time: Mar 26 03:26:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

