

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086114.D
 Acq On : 25 Mar 2025 20:40
 Operator : JC\MD
 Sample : Q1632-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PIER-1-2

Quant Time: Mar 26 03:27:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	308693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128081	57.309	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.620%
35) Dibromofluoromethane	8.165	113	77245	35.851	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	71.700%#
50) Toluene-d8	10.565	98	382552	48.921	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.847	95	125326	44.939	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.880%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	303581	444.620	ug/l	100
20) Methylene Chloride	5.289	84	4582	2.294	ug/l	92
25) 2-Butanone	7.483	43	12418	12.604	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086114.D
Acq On : 25 Mar 2025 20:40
Operator : JC\MD
Sample : Q1632-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PIER-1-2

Quant Time: Mar 26 03:27:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

