

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
 Data File : VN086115.D
 Acq On : 25 Mar 2025 21:04
 Operator : JC\MD
 Sample : Q1636-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Mar 26 03:27:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321904	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134485	56.766	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.540%	
35) Dibromofluoromethane	8.165	113	114986	51.177	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.360%	
50) Toluene-d8	10.565	98	422820	51.851	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.700%	
62) 4-Bromofluorobenzene	12.847	95	132958	45.719	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	90782	125.427	ug/l	99
20) Methylene Chloride	5.277	84	5090	2.404	ug/l #	93
25) 2-Butanone	7.483	43	11789	11.288	ug/l #	85
43) Isopropyl Acetate	8.688	43	29222	5.403	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032525\
Data File : VN086115.D
Acq On : 25 Mar 2025 21:04
Operator : JC\MD
Sample : Q1636-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Mar 26 03:27:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

