

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
 Data File : VN054690.D
 Acq On : 26 Mar 2019 11:28
 Operator : JC/SP
 Sample : K2056-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-8

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 9:10:14 AM

Quant Time: Mar 27 04:23:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	560042	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	884762	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	747953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	255645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	361837	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
35) Dibromofluoromethane	7.59	113	283551	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
50) Toluene-d8	10.09	98	1069318	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	12.40	95	319783	43.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.70%	
Target Compounds						
16) Acetone	3.82	43	11124	5.311	ug/l #	77
20) Methylene Chloride	4.55	84	14169m	2.295	ug/l	
43) Isopropyl Acetate	8.17	43	13360	1.344	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032619\
Data File : VN054690.D
Acq On : 26 Mar 2019 11:28
Operator : JC/SP
Sample : K2056-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-8

Manual Integrations
APPROVED

MMDadoda
3/27/2019 9:10:14 AM

Quant Time: Mar 27 04:23:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

