

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
 Data File : VN054696.D
 Acq On : 26 Mar 2019 14:06
 Operator : JC/SP
 Sample : K2081-23
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S4C(12)

Quant Time: Mar 27 04:39:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	527777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	875270	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	746468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	271523	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	353732	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
35) Dibromofluoromethane	7.59	113	272745	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
50) Toluene-d8	10.09	98	1043301	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
62) 4-Bromofluorobenzene	12.40	95	327622	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
Target Compounds						
16) Acetone	3.82	43	10561	5.351	ug/l #	82
20) Methylene Chloride	4.55	84	18196	3.128	ug/l #	76
43) Isopropyl Acetate	8.17	43	15168	1.543	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032619\
Data File : VN054696.D
Acq On : 26 Mar 2019 14:06
Operator : JC/SP
Sample : K2081-23
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S4C(12)

Quant Time: Mar 27 04:39:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

