

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
 Data File : VN054698.D
 Acq On : 26 Mar 2019 14:58
 Operator : JC/SP
 Sample : K2081-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S3A(22)

Quant Time: Mar 27 04:43:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	550333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	883108	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	752279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	271882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	354519	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
35) Dibromofluoromethane	7.59	113	280253	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	10.09	98	1059954	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.40	95	333396	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
Target Compounds						
16) Acetone	3.83	43	12473	6.061	ug/l #	79
20) Methylene Chloride	4.55	84	17874	2.946	ug/l	91
43) Isopropyl Acetate	8.17	43	15648	1.578	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032619\
Data File : VN054698.D
Acq On : 26 Mar 2019 14:58
Operator : JC/SP
Sample : K2081-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S3A(22)

Quant Time: Mar 27 04:43:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

