

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
 Data File : VN054701.D
 Acq On : 26 Mar 2019 16:17
 Operator : JC/SP
 Sample : K1693-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-032519-A

Quant Time: Mar 27 04:56:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	558457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	883534	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	740889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	275931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	360504	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
35) Dibromofluoromethane	7.59	113	276432	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
50) Toluene-d8	10.09	98	1054266	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.40	95	326053	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
Target Compounds						
16) Acetone	3.82	43	13781	6.599	ug/l	92
20) Methylene Chloride	4.55	84	2120270	344.411	ug/l #	82
43) Isopropyl Acetate	8.17	43	14945	1.506	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
Data File : VN054701.D
Acq On : 26 Mar 2019 16:17
Operator : JC/SP
Sample : K1693-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-01-032519-A

Quant Time: Mar 27 04:56:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

