

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032619\
 Data File : VN054704.D
 Acq On : 26 Mar 2019 18:41
 Operator : JC/SP
 Sample : K2106-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-1

Quant Time: Mar 27 05:09:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	555411	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	892844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	739358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	254247	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	377313	54.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.24%	
35) Dibromofluoromethane	7.59	113	288405	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
50) Toluene-d8	10.09	98	1058524	48.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.90%	
62) 4-Bromofluorobenzene	12.40	95	315694	42.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032619\
Data File : VN054704.D
Acq On : 26 Mar 2019 18:41
Operator : JC/SP
Sample : K2106-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-1

Quant Time: Mar 27 05:09:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

