

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032619\
 Data File : VN054706.D
 Acq On : 26 Mar 2019 19:33
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 05:14:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.478	0.492	-2.9	79	0.00
3 P	Chloromethane	0.650	0.698	-7.4	83	0.00
4 C	Vinyl Chloride	0.653	0.638	2.3#	74	0.00
5 T	Bromomethane	0.422	0.363	14.0	71	0.00
6 T	Chloroethane	0.392	0.358	8.7	72	0.00
7 T	Trichlorofluoromethane	0.793	0.800	-0.9	78	0.00
8 T	Diethyl Ether	0.305	0.264	13.4	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.466	0.412	11.6	71	0.00
10 T	Methyl Iodide	0.686	0.619	9.8	68	0.00
11 T	Tert butyl alcohol	0.038	0.038	0.0	73	0.00
12 CM	1,1-Dichloroethene	0.482	0.418	13.3#	68	0.00
13 T	Acrolein	0.052	0.044	15.4	64	0.00
14 T	Allyl chloride	0.908	1.004	-10.6	86	0.00
15 T	Acrylonitrile	0.188	0.209	-11.2	83	0.00
16 T	Acetone	0.187	0.162	13.4	64	0.00
17 T	Carbon Disulfide	1.292	1.265	2.1	79	0.00
18 T	Methyl Acetate	0.417	0.476	-14.1	87	0.00
19 T	Methyl tert-butyl Ether	1.407	1.455	-3.4	78	0.00
20 T	Methylene Chloride	0.551	0.572	-3.8	83	0.00
21 T	trans-1,2-Dichloroethene	0.492	0.535	-8.7	83	0.00
22 T	Diisopropyl ether	1.791	2.060	-15.0	88	0.00
23 T	Vinyl Acetate	1.265	1.484	-17.3	85	0.00
24 P	1,1-Dichloroethane	0.964	1.080	-12.0	86	0.00
25 T	2-Butanone	0.240	0.262	-9.2	77	0.00
26 T	2,2-Dichloropropane	0.688	0.735	-6.8	80	0.00
27 T	cis-1,2-Dichloroethene	0.587	0.621	-5.8	82	0.00
28 T	Bromochloromethane	0.475	0.553	-16.4	88	0.00
29 T	Tetrahydrofuran	0.155	0.178	-14.8	82	0.00
30 C	Chloroform	0.941	1.044	-10.9#	85	0.00
31 T	Cyclohexane	1.121	1.003	10.5	84	0.00
32 T	1,1,1-Trichloroethane	0.777	0.869	-11.8	84	0.00
33 S	1,2-Dichloroethane-d4	0.622	0.683	-9.8	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.322	0.346	-7.5	81	0.00
36 T	1,1-Dichloropropene	0.468	0.491	-4.9	83	0.00
37 T	Ethyl Acetate	0.333	0.365	-9.6	81	0.00
38 T	Carbon Tetrachloride	0.454	0.476	-4.8	83	0.00
39 T	Methylcyclohexane	0.535	0.519	3.0	76	0.00
40 TM	Benzene	1.337	1.451	-8.5	84	0.00
41 T	Methacrylonitrile	0.203	0.209	-3.0	81	0.00
42 TM	1,2-Dichloroethane	0.487	0.522	-7.2	84	0.00
43 T	Isopropyl Acetate	0.562	0.634	-12.8	83	0.00
44 TM	Trichloroethene	0.372	0.374	-0.5	79	0.00
45 C	1,2-Dichloropropane	0.359	0.415	-15.6#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032619\
 Data File : VN054706.D
 Acq On : 26 Mar 2019 19:33
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 05:14:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.230	-7.5	82	0.00
47 T	Bromodichloromethane	0.444	0.506	-14.0	86	0.00
48 T	Methyl methacrylate	0.305	0.334	-9.5	83	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	74	0.00
50 S	Toluene-d8	1.211	1.285	-6.1	81	0.00
51 T	4-Methyl-2-Pentanone	0.311	0.355	-14.1	82	0.00
52 CM	Toluene	0.822	0.869	-5.7#	82	0.00
53 T	t-1,3-Dichloropropene	0.414	0.484	-16.9	83	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.575	-12.7	83	0.00
55 T	1,1,2-Trichloroethane	0.296	0.322	-8.8	82	0.00
56 T	Ethyl methacrylate	0.397	0.438	-10.3	80	0.00
57 T	1,3-Dichloropropane	0.514	0.567	-10.3	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.159	2.5	71	0.00
59 T	2-Hexanone	0.203	0.222	-9.4	80	0.00
60 T	Dibromochloromethane	0.315	0.349	-10.8	81	0.00
61 T	1,2-Dibromoethane	0.292	0.311	-6.5	79	0.00
62 S	4-Bromofluorobenzene	0.412	0.435	-5.6	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.415	0.399	3.9	77	0.00
65 PM	Chlorobenzene	1.017	1.047	-2.9	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.394	-7.9	82	0.00
67 C	Ethyl Benzene	1.796	1.881	-4.7#	83	0.00
68 T	m/p-Xylenes	0.680	0.693	-1.9	80	0.00
69 T	o-Xylene	0.664	0.679	-2.3	81	0.00
70 T	Styrene	1.079	1.134	-5.1	80	0.00
71 P	Bromoform	0.222	0.244	-9.9	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
73 T	Isopropylbenzene	4.029	4.052	-0.6	81	0.00
74 T	N-amyl acetate	1.159	1.312	-13.2	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.920	1.004	-9.1	82	0.00
76 T	1,2,3-Trichloropropane	0.793	0.848	-6.9	81	0.00
77 T	Bromobenzene	0.999	1.006	-0.7	79	0.00
78 T	n-propylbenzene	4.369	4.482	-2.6	81	0.00
79 T	2-Chlorotoluene	2.724	2.776	-1.9	82	0.00
80 T	1,3,5-Trimethylbenzene	3.302	3.339	-1.1	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.232	-13.2	81	0.00
82 T	4-Chlorotoluene	2.641	2.732	-3.4	80	0.00
83 T	tert-Butylbenzene	2.874	2.838	1.3	78	0.00
84 T	1,2,4-Trimethylbenzene	3.240	3.263	-0.7	77	0.00
85 T	sec-Butylbenzene	3.583	3.571	0.3	79	0.00
86 T	p-Isopropyltoluene	3.206	3.136	2.2	75	0.00
87 T	1,3-Dichlorobenzene	1.656	1.673	-1.0	77	0.00
88 T	1,4-Dichlorobenzene	1.611	1.604	0.4	75	0.00
89 T	n-Butylbenzene	2.507	2.529	-0.9	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032619\
Data File : VN054706.D
Acq On : 26 Mar 2019 19:33
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 27 05:14:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.498	0.539	-8.2	81	0.00
91 T	1,2-Dichlorobenzene	1.604	1.646	-2.6	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.135	0.146	-8.1	73	0.00
93 T	1,2,4-Trichlorobenzene	0.800	0.785	1.9	69	0.00
94 T	Hexachlorobutadiene	0.523	0.456	12.8	71	0.00
95 T	Naphthalene	1.783	1.753	1.7	68	0.00
96 T	1,2,3-Trichlorobenzene	0.796	0.776	2.5	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6