


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032619\
Data File : VN054684.D
Acq On : 26 Mar 2019 8:40
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Title : SW846 8260
Last Update : Fri Mar 22 02:09:07 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	14201	PASS
75	95	30	60	51.6	33912	PASS
95	95	100	100	100.0	65728	PASS
96	95	5	9	7.2	4752	PASS
173	174	0.00	2	0.6	343	PASS
174	95	50	100	83.7	54992	PASS
175	174	5	9	7.6	4192	PASS
176	174	95	101	96.6	53136	PASS
177	176	5	9	7.4	3906	PASS