

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\  
 Data File : VN060715.D  
 Acq On : 26 Mar 2020 15:26  
 Operator : JC/MD  
 Sample : L1997-05  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW203D2-20200315

Quant Time: Mar 27 07:06:22 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N031820W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 18 08:49:09 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	330594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308948	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142709	50.00	ug/l	0.00

## System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.00	65	130014	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
35) Dibromofluoromethane	7.57	113	98718	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
50) Toluene-d8	10.08	98	405366	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.40	95	152294	50.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.78%	

## Target Compounds

					Qvalue	
24) 1,1-Dichloroethane	5.82	63	8257	1.925	ug/l	# 94
27) cis-1,2-Dichloroethene	6.79	96	1340	0.522	ug/l	84
44) Trichloroethene	8.82	130	1409	0.578	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\  
 Data File : VN060715.D  
 Acq On : 26 Mar 2020 15:26  
 Operator : JC/MD  
 Sample : L1997-05  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW203D2-20200315

Quant Time: Mar 27 07:06:22 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N031820W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 18 08:49:09 2020  
 Response via : Initial Calibration

