

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\
 Data File : VN060718.D
 Acq On : 26 Mar 2020 16:58
 Operator : JC/MD
 Sample : PB127792TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127792TB

Quant Time: Mar 27 07:15:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	338830	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	320724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149831	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131893	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
35) Dibromofluoromethane	7.56	113	100560	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
50) Toluene-d8	10.08	98	421205	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	12.40	95	157870	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
16) Acetone	3.78	43	7272	7.375	ug/l	95
20) Methylene Chloride	4.52	84	5134	2.117	ug/l	93
43) Isopropyl Acetate	8.14	43	39095	6.687	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	3793	1.140	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032620\
Data File : VN060718.D
Acq On : 26 Mar 2020 16:58
Operator : JC/MD
Sample : PB127792TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127792TB

Quant Time: Mar 27 07:15:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

