

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\
 Data File : VN060723.D
 Acq On : 26 Mar 2020 18:52
 Operator : JC/MD
 Sample : PB127792ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127792ZHE#05

Quant Time: Mar 27 07:29:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	207969	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	349063	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	329887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	154343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	135905	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
35) Dibromofluoromethane	7.56	113	103981	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	10.08	98	431864	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.40	95	164435	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
Target Compounds						
16) Acetone	3.79	43	6260	6.190	ug/l #	86
20) Methylene Chloride	4.52	84	4305	1.731	ug/l	95
43) Isopropyl Acetate	8.14	43	60552	10.053	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032620\
Data File : VN060723.D
Acq On : 26 Mar 2020 18:52
Operator : JC/MD
Sample : PB127792ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127792ZHE#05

Quant Time: Mar 27 07:29:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

