

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\
 Data File : VN060724.D
 Acq On : 26 Mar 2020 19:15
 Operator : JC/MD
 Sample : PB127792ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127792ZHE#01

Quant Time: Mar 27 07:31:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	328744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	307378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141859	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	129342	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
35) Dibromofluoromethane	7.57	113	98258	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
50) Toluene-d8	10.08	98	404696	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.40	95	152386	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
Target Compounds						
16) Acetone	3.78	43	6414	6.758	ug/l	99
20) Methylene Chloride	4.52	84	2974	1.274	ug/l	92
43) Isopropyl Acetate	8.14	43	62193	10.964	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032620\
Data File : VN060724.D
Acq On : 26 Mar 2020 19:15
Operator : JC/MD
Sample : PB127792ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127792ZHE#01

Quant Time: Mar 27 07:31:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

