

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032620\
 Data File : VN060729.D
 Acq On : 26 Mar 2020 21:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 07:44:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.484	0.465	3.9	82	0.00
3 P	Chloromethane	0.904	0.787	12.9	78	0.00
4 C	Vinyl Chloride	0.707	0.624	11.7#	79	0.00
5 T	Bromomethane	0.305	0.296	3.0	87	0.00
6 T	Chloroethane	0.407	0.391	3.9	83	0.00
7 T	Trichlorofluoromethane	0.871	0.836	4.0	85	0.00
8 T	Diethyl Ether	0.350	0.343	2.0	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.505	-0.2	88	0.00
10 T	Methyl Iodide	0.449	0.427	4.9	89	0.00
11 T	Tert butyl alcohol	0.101	0.088	12.9	80	0.00
12 CM	1,1-Dichloroethene	0.541	0.516	4.6#	86	0.00
13 T	Acrolein	0.112	0.116	-3.6	94	0.00
14 T	Allyl chloride	0.952	0.895	6.0	87	0.00
15 T	Acrylonitrile	0.297	0.289	2.7	86	0.00
16 T	Acetone	0.243	0.228	6.2	84	0.00
17 T	Carbon Disulfide	1.646	1.491	9.4	81	0.00
18 T	Methyl Acetate	0.684	0.666	2.6	85	0.00
19 T	Methyl tert-butyl Ether	1.861	1.772	4.8	86	0.00
20 T	Methylene Chloride	0.598	0.578	3.3	86	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.544	5.7	85	0.00
22 T	Diisopropyl ether	2.012	2.037	-1.2	89	0.00
23 T	Vinyl Acetate	1.641	1.661	-1.2	88	0.00
24 P	1,1-Dichloroethane	1.097	1.078	1.7	87	0.00
25 T	2-Butanone	0.393	0.375	4.6	84	0.00
26 T	2,2-Dichloropropane	0.863	0.829	3.9	87	0.00
27 T	cis-1,2-Dichloroethene	0.656	0.631	3.8	87	0.00
28 T	Bromochloromethane	0.525	0.507	3.4	85	0.00
29 T	Tetrahydrofuran	0.278	0.260	6.5	84	0.00
30 C	Chloroform	1.051	1.024	2.6#	87	0.00
31 T	Cyclohexane	1.106	1.023	7.5	86	0.00
32 T	1,1,1-Trichloroethane	0.906	0.892	1.5	86	0.00
33 S	1,2-Dichloroethane-d4	0.686	0.675	1.6	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.302	0.308	-2.0	89	0.00
36 T	1,1-Dichloropropene	0.484	0.489	-1.0	87	0.00
37 T	Ethyl Acetate	0.523	0.496	5.2	84	0.00
38 T	Carbon Tetrachloride	0.434	0.432	0.5	84	0.00
39 T	Methylcyclohexane	0.575	0.593	-3.1	88	0.00
40 TM	Benzene	1.462	1.450	0.8	87	0.00
41 T	Methacrylonitrile	0.223	0.230	-3.1	90	0.00
42 TM	1,2-Dichloroethane	0.499	0.501	-0.4	87	0.00
43 T	Isopropyl Acetate	0.863	0.859	0.5	87	0.00
44 TM	Trichloroethene	0.368	0.366	0.5	86	0.00
45 C	1,2-Dichloropropane	0.386	0.398	-3.1#	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032620\
 Data File : VN060729.D
 Acq On : 26 Mar 2020 21:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 07:44:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.233	0.236	-1.3	86	0.00
47 T	Bromodichloromethane	0.484	0.493	-1.9	88	0.00
48 T	Methyl methacrylate	0.384	0.380	1.0	86	0.00
49 T	1,4-Dioxane	0.005	0.004	20.0	77	0.00
50 S	Toluene-d8	1.235	1.220	1.2	85	0.00
51 T	4-Methyl-2-Pentanone	0.491	0.492	-0.2	85	0.00
52 CM	Toluene	0.883	0.884	-0.1#	86	0.00
53 T	t-1,3-Dichloropropene	0.557	0.570	-2.3	87	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.612	-0.2	87	0.00
55 T	1,1,2-Trichloroethane	0.340	0.341	-0.3	87	0.00
56 T	Ethyl methacrylate	0.529	0.545	-3.0	85	0.00
57 T	1,3-Dichloropropane	0.589	0.590	-0.2	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.263	0.259	1.5	81	0.00
59 T	2-Hexanone	0.354	0.353	0.3	83	0.00
60 T	Dibromochloromethane	0.351	0.358	-2.0	86	0.00
61 T	1,2-Dibromoethane	0.341	0.344	-0.9	85	0.00
62 S	4-Bromofluorobenzene	0.453	0.456	-0.7	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.393	0.375	4.6	84	0.00
65 PM	Chlorobenzene	0.991	0.970	2.1	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.358	0.3	87	0.00
67 C	Ethyl Benzene	1.843	1.848	-0.3#	86	0.00
68 T	m/p-Xylenes	0.685	0.681	0.6	86	0.00
69 T	o-Xylene	0.654	0.651	0.5	85	0.00
70 T	Styrene	1.069	1.089	-1.9	86	0.00
71 P	Bromoform	0.268	0.277	-3.4	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.752	3.586	4.4	87	0.00
74 T	N-amyl acetate	1.734	1.662	4.2	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.135	1.078	5.0	87	0.00
76 T	1,2,3-Trichloropropane	1.079	1.063	1.5	94	0.00
77 T	Bromobenzene	0.904	0.849	6.1	87	0.00
78 T	n-propylbenzene	4.433	4.288	3.3	88	0.00
79 T	2-Chlorotoluene	2.649	2.475	6.6	86	0.00
80 T	1,3,5-Trimethylbenzene	3.184	3.049	4.2	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.420	0.383	8.8	83	0.00
82 T	4-Chlorotoluene	2.698	2.577	4.5	87	0.00
83 T	tert-Butylbenzene	2.640	2.544	3.6	88	0.00
84 T	1,2,4-Trimethylbenzene	3.150	3.035	3.7	87	0.00
85 T	sec-Butylbenzene	3.550	3.543	0.2	90	0.00
86 T	p-Isopropyltoluene	3.192	3.166	0.8	90	0.00
87 T	1,3-Dichlorobenzene	1.595	1.554	2.6	89	0.00
88 T	1,4-Dichlorobenzene	1.602	1.554	3.0	89	0.00
89 T	n-Butylbenzene	2.882	3.006	-4.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032620\
Data File : VN060729.D
Acq On : 26 Mar 2020 21:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 27 07:44:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.464	0.472	-1.7	89	0.00
91 T	1,2-Dichlorobenzene	1.525	1.490	2.3	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.255	0.214	16.1	84	0.00
93 T	1,2,4-Trichlorobenzene	0.936	0.973	-4.0	91	0.00
94 T	Hexachlorobutadiene	0.459	0.475	-3.5	95	0.00
95 T	Naphthalene	2.816	2.759	2.0	86	0.00
96 T	1,2,3-Trichlorobenzene	0.922	0.917	0.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6