

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032620\
 Data File : VN060707.D
 Acq On : 26 Mar 2020 12:06
 Operator : JC/MD
 Sample : VN0326WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0326WBL01

Quant Time: Mar 27 05:45:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	189341	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	313940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	282983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	115708	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	122667	47.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.38%	
35) Dibromofluoromethane	7.56	113	94318	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
50) Toluene-d8	10.08	98	386749	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.40	95	134943	47.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.98%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032620\
Data File : VN060707.D
Acq On : 26 Mar 2020 12:06
Operator : JC/MD
Sample : VN0326WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0326WBL01

Quant Time: Mar 27 05:45:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1824

Delta R.T. -0.01 min

Lab File: VN060707.D

Acq: 26 Mar 2020 12:06

Instrument :

MSVOA_N

ClientSampleId :


VN0326WBL01

Tot Ion:168 Resp: 189341

Ion Ratio Lower Upper




168 100

99 57.2 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 47.194 ug/l

RT: 8.00 min Scan# 1937

Delta R.T. -0.00 min

Lab File: VN060707.D


Acq: 26 Mar 2020 12:06

Tgt Ion: 65 Resp: 122667

Ion Ratio Lower Upper



65 100


67 52.9 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.56 min Scan# 2112
 Delta R.T. -0.00 min
 Lab File: VN060707.D
 Acq: 26 Mar 2020 12:06




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0326WBL01


Tot Ion	Ratio	Lower	Upper
114	100		
63	22.6	0.0	43.8
88	16.9	0.0	33.0



#35
 Dibromofluoromethane
 Concen: 49.686 ug/l
 RT: 7.56 min Scan# 1800
 Delta R.T. -0.00 min
 Lab File: VN060707.D
 Acq: 26 Mar 2020 12:06

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.2	82.9	124.3
192	19.8	15.8	23.8





#50

Toluene-d8

Concen: 49.879 ug/l

RT: 10.08 min Scan# 2582

Delta R.T. -0.00 min

Lab File: VN060707.D

Acq: 26 Mar 2020 12:06

Instrument :

MSVOA_N

ClientSampleId :





VN0326WBL01

Tot Ion: 98 Resp: 386749

Ion Ratio Lower Upper

98 100

100 65.3 51.8 77.8



#62

4-Bromofluorobenzene

Concen: 47.487 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.00 min

Lab File: VN060707.D

Acq: 26 Mar 2020 12:06




Tgt Ion: 95 Resp: 134943


Ion Ratio Lower Upper

95 100

174 78.7 0.0 157.6

176 75.4 0.0 150.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2993
Delta R.T. -0.00 min
Lab File: VN060707.D
Acq: 26 Mar 2020 12:06

Instrument :
MSVOA_N
ClientSampleId :
VN0326WBL01

Tot Ion:117 Resp: 282983
Ion Ratio Lower Upper
117 100
82 60.0 46.7 70.1
119 32.7 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3597
Delta R.T. -0.00 min
Lab File: VN060707.D
Acq: 26 Mar 2020 12:06

Tgt Ion:152 Resp: 115708
Ion Ratio Lower Upper
152 100
115 61.1 30.2 90.6
150 159.1 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

