

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066351.D
 Acq On : 26 Mar 2021 15:59
 Operator : JC/MD
 Sample : M1788-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW201D1-20210320

Quant Time: Mar 27 01:49:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	501760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	757368	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	628310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	209010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	356849	53.50	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	= 107.00%		
35) Dibromofluoromethane	7.507	113	276952	55.16	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	= 110.32%		
50) Toluene-d8	10.028	98	1068788	56.12	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	= 112.24%#		
62) 4-Bromofluorobenzene	12.345	95	299950	45.35	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 90.70%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.818	85	41605	6.58	ug/l	99
7) Trichlorofluoromethane	2.963	101	1233293	123.99	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.680	101	20051	3.57	ug/l	98
12) 1,1-Dichloroethene	3.661	96	220093	38.27	ug/l	99
24) 1,1-Dichloroethane	5.742	63	33451	3.41	ug/l	98
27) cis-1,2-Dichloroethene	6.726	96	16895	2.78	ug/l	89
30) Chloroform	7.287	83	36808	3.76	ug/l	99
32) 1,1,1-Trichloroethane	7.480	97	45228	5.26	ug/l	98
44) Trichloroethene	8.765	130	670342	118.60	ug/l	99
55) 1,1,2-Trichloroethane	10.502	97	2020	0.38	ug/l	93
64) Tetrachloroethene	10.572	164	631475	118.09	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
Data File : VN066351.D
Acq On : 26 Mar 2021 15:59
Operator : JC/MD
Sample : M1788-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW201D1-20210320

Quant Time: Mar 27 01:49:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

