

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066354.D
 Acq On : 26 Mar 2021 17:09
 Operator : JC/MD
 Sample : M1788-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE105D1-20210322

Quant Time: Mar 27 01:49:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	494542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	742890	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	646343	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	226821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	347161	52.81	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.62%	
35) Dibromofluoromethane	7.509	113	272224	55.28	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.56%	
50) Toluene-d8	10.028	98	1044462	55.91	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 111.82%	
62) 4-Bromofluorobenzene	12.345	95	310619	47.87	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.74%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	3.679	101	19791	3.58	ug/l	96
12) 1,1-Dichloroethene	3.658	96	2678	0.47	ug/l	93
27) cis-1,2-Dichloroethene	6.734	96	5146	0.86	ug/l	76
44) Trichloroethene	8.764	130	491293	88.62	ug/l	99
64) Tetrachloroethene	10.572	164	1441	0.26	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
Data File : VN066354.D
Acq On : 26 Mar 2021 17:09
Operator : JC/MD
Sample : M1788-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE105D1-20210322

Quant Time: Mar 27 01:49:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

