

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066355.D
 Acq On : 26 Mar 2021 17:33
 Operator : JC/MD
 Sample : M1788-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE108D1-20210322

Quant Time: Mar 27 01:49:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	521182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	773784	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	687288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	222956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	358111	51.69	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.38%		
35) Dibromofluoromethane	7.507	113	278968	54.38	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.76%		
50) Toluene-d8	10.028	98	1121103	57.62	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 115.24%#		
62) 4-Bromofluorobenzene	12.345	95	302911	44.82	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 89.64%		
Target Compounds						
44) Trichloroethene	8.764	130	199944	34.62	ug/l	99
64) Tetrachloroethene	10.572	164	7128	1.22	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
Data File : VN066355.D
Acq On : 26 Mar 2021 17:33
Operator : JC/MD
Sample : M1788-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE108D1-20210322

Quant Time: Mar 27 01:49:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

