

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066361.D
 Acq On : 26 Mar 2021 19:53
 Operator : JC/MD
 Sample : M1788-07 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE108D2-20210322

Quant Time: Mar 27 01:51:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	509427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	752058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	640345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	238751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	334689	49.42	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.84%
35) Dibromofluoromethane	7.507	113	269250	54.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.02%
50) Toluene-d8	10.028	98	1020797	53.98	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.96%
62) 4-Bromofluorobenzene	12.345	95	298289	45.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.82%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	6.734	96	3728	0.60	ug/l	79
44) Trichloroethene	8.765	130	1484328	264.47	ug/l	100
64) Tetrachloroethene	10.564	164	1628	0.30	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
Data File : VN066361.D
Acq On : 26 Mar 2021 19:53
Operator : JC/MD
Sample : M1788-07 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE108D2-20210322

Quant Time: Mar 27 01:51:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

