

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066341.D
 Acq On : 26 Mar 2021 11:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 01:46:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.630	0.661	-4.9	97	0.00
3 P	Chloromethane	0.739	0.711	3.8	100	0.00
4 C	Vinyl Chloride	0.721	0.743	-3.1#	97	0.00
5 T	Bromomethane	0.483	0.505	-4.6	97	0.00
6 T	Chloroethane	0.449	0.458	-2.0	97	0.00
7 T	Trichlorofluoromethane	0.991	1.035	-4.4	97	0.00
8 T	Diethyl Ether	0.404	0.406	-0.5	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.560	0.579	-3.4	97	0.00
10 T	Methyl Iodide	0.794	0.795	-0.1	90	0.00
11 T	Tert butyl alcohol	0.101	0.099	2.0	85	0.01
12 CM	1,1-Dichloroethene	0.573	0.578	-0.9#	95	0.00
13 T	Acrolein	0.066	0.057	13.6	98	0.00
14 T	Allyl chloride	1.133	1.180	-4.1	94	0.00
15 T	Acrylonitrile	0.283	0.286	-1.1	92	0.00
16 T	Acetone	0.359	0.349	2.8	93	0.00
17 T	Carbon Disulfide	1.692	1.716	-1.4	97	0.00
18 T	Methyl Acetate	0.938	0.961	-2.5	94	0.00
19 T	Methyl tert-butyl Ether	1.699	1.690	0.5	94	0.00
20 T	Methylene Chloride	0.692	0.659	4.8	94	0.00
21 T	trans-1,2-Dichloroethene	0.520	0.519	0.2	96	0.00
22 T	Diisopropyl ether	1.768	1.797	-1.6	96	0.00
23 T	Vinyl Acetate	1.505	1.556	-3.4	94	0.00
24 P	1,1-Dichloroethane	0.976	0.980	-0.4	95	0.00
25 T	2-Butanone	0.393	0.386	1.8	89	0.00
26 T	2,2-Dichloropropane	0.870	0.872	-0.2	94	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.603	0.5	95	0.00
28 T	Bromochloromethane	0.492	0.473	3.9	90	0.00
29 T	Tetrahydrofuran	0.256	0.256	0.0	89	0.00
30 C	Chloroform	0.976	0.990	-1.4#	95	0.00
31 T	Cyclohexane	0.920	0.890	3.3	94	0.00
32 T	1,1,1-Trichloroethane	0.857	0.858	-0.1	93	0.00
33 S	1,2-Dichloroethane-d4	0.665	0.649	2.4	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.331	0.328	0.9	90	0.00
36 T	1,1-Dichloropropene	0.470	0.482	-2.6	95	0.00
37 T	Ethyl Acetate	0.517	0.533	-3.1	91	0.00
38 T	Carbon Tetrachloride	0.472	0.481	-1.9	93	0.00
39 T	Methylcyclohexane	0.573	0.594	-3.7	94	0.00
40 TM	Benzene	1.419	1.475	-3.9	95	0.00
41 T	Methacrylonitrile	0.257	0.253	1.6	95	0.00
42 TM	1,2-Dichloroethane	0.508	0.521	-2.6	94	0.00
43 T	Isopropyl Acetate	0.884	0.914	-3.4	93	0.00
44 TM	Trichloroethene	0.373	0.381	-2.1	94	0.00
45 C	1,2-Dichloropropane	0.378	0.389	-2.9#	95	0.00
46 T	Dibromomethane	0.248	0.258	-4.0	94	0.00
47 T	Bromodichloromethane	0.512	0.533	-4.1	95	0.00
48 T	Methyl methacrylate	0.420	0.435	-3.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
 Data File : VN066341.D
 Acq On : 26 Mar 2021 11:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 01:46:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	94	0.00
50 S	Toluene-d8	1.257	1.266	-0.7	91	0.00
51 T	4-Methyl-2-Pentanone	0.516	0.536	-3.9	90	0.00
52 CM	Toluene	0.871	0.920	-5.6#	95	0.00
53 T	t-1,3-Dichloropropene	0.556	0.590	-6.1	96	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.634	-6.2	95	0.00
55 T	1,1,2-Trichloroethane	0.350	0.365	-4.3	94	0.00
56 T	Ethyl methacrylate	0.537	0.567	-5.6	93	0.00
57 T	1,3-Dichloropropane	0.581	0.609	-4.8	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.048	0.039	18.8	71	0.00
59 T	2-Hexanone	0.362	0.392	-8.3	91	0.00
60 T	Dibromochloromethane	0.389	0.420	-8.0	95	0.00
61 T	1,2-Dibromoethane	0.365	0.382	-4.7	93	0.00
62 S	4-Bromofluorobenzene	0.437	0.442	-1.1	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.426	0.438	-2.8	94	0.00
65 PM	Chlorobenzene	0.997	1.053	-5.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.415	-8.1	95	0.00
67 C	Ethyl Benzene	1.809	1.946	-7.6#	95	0.00
68 T	m/p-Xylenes	0.688	0.726	-5.5	93	0.00
69 T	o-Xylene	0.663	0.705	-6.3	94	0.00
70 T	Styrene	1.076	1.168	-8.6	92	0.00
71 P	Bromoform	0.319	0.351	-10.0	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.719	3.646	2.0	93	0.00
74 T	N-amyl acetate	1.307	1.453	-11.2	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.169	1.3	96	0.00
76 T	1,2,3-Trichloropropane	1.097	1.114	-1.5	97	0.00
77 T	Bromobenzene	0.939	0.915	2.6	93	0.00
78 T	n-propylbenzene	4.044	4.105	-1.5	93	0.00
79 T	2-Chlorotoluene	2.487	2.421	2.7	93	0.00
80 T	1,3,5-Trimethylbenzene	3.042	3.046	-0.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.365	0.368	-0.8	88	0.00
82 T	4-Chlorotoluene	2.475	2.495	-0.8	92	0.00
83 T	tert-Butylbenzene	2.619	2.560	2.3	93	0.00
84 T	1,2,4-Trimethylbenzene	2.956	2.994	-1.3	94	0.00
85 T	sec-Butylbenzene	3.393	3.434	-1.2	92	0.00
86 T	p-Isopropyltoluene	3.039	3.099	-2.0	92	0.00
87 T	1,3-Dichlorobenzene	1.576	1.612	-2.3	94	0.00
88 T	1,4-Dichlorobenzene	1.636	1.667	-1.9	95	0.00
89 T	n-Butylbenzene	2.510	2.631	-4.8	93	0.00
90 T	Hexachloroethane	0.548	0.551	-0.5	94	0.00
91 T	1,2-Dichlorobenzene	1.549	1.623	-4.8	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.246	-8.4	97	0.00
93 T	1,2,4-Trichlorobenzene	0.961	1.023	-6.5	98	0.00
94 T	Hexachlorobutadiene	0.524	0.533	-1.7	96	0.00
95 T	Naphthalene	2.899	2.885	0.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032621\
Data File : VN066341.D
Acq On : 26 Mar 2021 11:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 27 01:46:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.966	0.984	-1.9	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6