

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071583.D
 Acq On : 26 Mar 2022 19:26
 Operator : JC\MD
 Sample : PB143590TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143590TB

Quant Time: Mar 28 02:20:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	826336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1393124	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1411886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	538758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	484946	47.236	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.480%
35) Dibromofluoromethane	8.022	113	394985	45.114	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.220%
50) Toluene-d8	10.439	98	1745080	49.415	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.727	95	606309	52.397	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.800%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.828	74	15213	2.719	ug/l	98
16) Acetone	4.292	43	26550	6.309	ug/l	95
20) Methylene Chloride	5.104	84	13643	1.391	ug/l #	86
43) Isopropyl Acetate	8.557	43	74863	3.021	ug/l #	88
95) Naphthalene	15.504	128	20416	3.911	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
Data File : VN071583.D
Acq On : 26 Mar 2022 19:26
Operator : JC\MD
Sample : PB143590TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143590TB

Quant Time: Mar 28 02:20:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

