

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071585.D
 Acq On : 26 Mar 2022 20:15
 Operator : JC\MD
 Sample : PB143590ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143590ZHE#13

Quant Time: Mar 28 02:21:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	762795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1270953	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1265082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	456041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	436201	46.028	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.060%
35) Dibromofluoromethane	8.016	113	357055	44.701	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.400%
50) Toluene-d8	10.439	98	1591280	49.391	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.780%
62) 4-Bromofluorobenzene	12.727	95	524709	49.704	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.400%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.828	74	14580	2.823	ug/l	89
16) Acetone	4.287	43	25847	6.654	ug/l	94
20) Methylene Chloride	5.087	84	14382	1.589	ug/l #	94
43) Isopropyl Acetate	8.563	43	138681	6.557	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
Data File : VN071585.D
Acq On : 26 Mar 2022 20:15
Operator : JC\MD
Sample : PB143590ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143590ZHE#13

Quant Time: Mar 28 02:21:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

