

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071587.D
 Acq On : 26 Mar 2022 21:03
 Operator : JC\MD
 Sample : PB143635TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143635TB

Quant Time: Mar 28 02:21:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	778504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1311778	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1300227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	475080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	448417	46.362	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.720%		
35) Dibromofluoromethane	8.016	113	370299	44.917	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	89.840%		
50) Toluene-d8	10.439	98	1634002	49.139	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.280%		
62) 4-Bromofluorobenzene	12.727	95	542069	49.751	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.500%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	22686	5.722	ug/l	90
20) Methylene Chloride	5.098	84	66077	7.153	ug/l	96
43) Isopropyl Acetate	8.563	43	152673	7.021	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
Data File : VN071587.D
Acq On : 26 Mar 2022 21:03
Operator : JC\MD
Sample : PB143635TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143635TB

Quant Time: Mar 28 02:21:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

