

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071588.D
 Acq On : 26 Mar 2022 21:27
 Operator : JC\MD
 Sample : N2140-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MHC

Quant Time: Mar 28 02:21:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	768337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1303119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1300220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	471919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	441887	46.291	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.580%
35) Dibromofluoromethane	8.022	113	366187	44.713	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.420%
50) Toluene-d8	10.439	98	1630241	49.351	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.727	95	547913	50.621	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.240%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	93474	23.890	ug/l	98
20) Methylene Chloride	5.098	84	63779	6.996	ug/l	96
25) 2-Butanone	7.339	43	13557	2.160	ug/l	97
43) Isopropyl Acetate	8.563	43	189200	8.860	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
Data File : VN071588.D
Acq On : 26 Mar 2022 21:27
Operator : JC\MD
Sample : N2140-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MHC

Quant Time: Mar 28 02:21:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

