

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071570.D
 Acq On : 26 Mar 2022 13:37
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 28 01:44:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Mar 28 01:40:30 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	95	0.00
2 M	Dichlorodifluoromethane	1.867	1.585	15.1	78	0.00
3 M	Chloromethane	2.201	1.842	16.3	80	0.00
4 M	Vinyl Chloride	2.524	2.133	15.5	82	0.00
5 M	Bromomethane	1.487	1.403	5.6	81	0.00
6 M	Chloroethane	1.616	1.384	14.4	83	0.00
7 M	Trichlorofluoromethane	2.943	2.526	14.2	82	-0.01
8 T	Diethyl Ether	1.236	1.039	15.9	81	0.00
9	1,1,2-Trichlorotrifluoroeth	1.762	1.542	12.5	82	0.00
10 M	1,1-Dichloroethene	1.738	1.450	16.6	80	0.00
11	Methyl Iodide	2.547	2.171	14.8	83	0.00
12	Methyl Acetate	2.449	2.172	11.3	86	0.00
13 M	Acrolein	0.261	0.222	14.9	86	0.00
14 M	Acrylonitrile	1.109	0.955	13.9	84	0.00
15 M	Acetone	0.319	0.263	17.6	76	0.00
16 M	Carbon Disulfide	4.525	3.455	23.6	75	0.00
17	Allyl chloride	2.819	2.314	17.9	81	0.00
18 M	Methylene Chloride	2.085	1.806	13.4	84	0.00
19 M	trans-1,2-Dichloroethene	1.859	1.557	16.2	82	0.00
20 T	Diisopropyl ether	5.898	5.044	14.5	81	0.00
21 M	1,1-Dichloroethane	3.561	3.017	15.3	82	0.00
22 M	cis-1,2-Dichloroethene	2.258	1.891	16.3	81	0.00
23 M	tert-Butyl Alcohol	0.396	0.348	12.1	89	0.00
24 M	Methyl tert-Butyl Ether	6.273	5.332	15.0	82	0.00
25 M	Chloroform	3.582	3.103	13.4	83	0.00
26	Cyclohexane	3.078	2.480	19.4	78	0.00
27 s	1,2-Dichloroethane-d4	2.114	2.145	-1.5	95	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
29	1,1-Dichloropropene	0.486	0.397	18.3	81	0.00
30 M	2-Butanone	0.270	0.222	17.8	80	0.00
31	2,2-Dichloropropane	0.553	0.443	19.9	79	0.00
32 M	1,1,1-Trichloroethane	0.566	0.476	15.9	83	0.00
33 M	Carbon Tetrachloride	0.479	0.385	19.6	79	0.00
34 M	Benzene	1.509	1.258	16.6	81	0.00
35	Methacrylonitrile	0.259	0.227	12.4	82	0.00
36 M	1,2-Dichloroethane	0.526	0.446	15.2	83	0.00
37 M	Trichloroethene	0.389	0.324	16.7	83	0.00
38	Methylcyclohexane	0.612	0.473	22.7	77	0.00
39 M	1,2-Dichloropropane	0.377	0.311	17.5	81	0.00
40	Dibromomethane	0.255	0.218	14.5	84	0.00
41 M	Bromodichloromethane	0.511	0.431	15.7	84	0.00
42 M	Vinyl Acetate	0.926	0.759	18.0	82	0.00
43	Ethyl Acetate	0.524	0.456	13.0	84	0.00
44	Isopropyl Acetate	0.841	0.720	14.4	87	0.00
45 T	1,4-Dioxane	0.008	0.008#	0.0	86	0.00
46	Methyl methacrylate	0.369	0.316	14.4	90	0.00
47	n-amyl Acetate	0.561	0.426	24.1	83	0.00
48 M	t-1,3-Dichloropropene	0.556	0.441	20.7	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032622\
 Data File : VN071570.D
 Acq On : 26 Mar 2022 13:37
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 28 01:44:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Mar 28 01:40:30 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.608	0.489	19.6	81	0.00
50 M	1,1,2-Trichloroethane	0.382	0.329	13.9	85	0.00
51	Ethyl methacrylate	0.590	0.468	20.7	83	0.00
52	1,3-Dichloropropane	0.652	0.541	17.0	81	0.00
53 M	Dibromochloromethane	0.395	0.325	17.7	83	0.00
54 M	1,2-Dibromoethane	0.393	0.328	16.5	84	0.00
55 M	2-Chloroethyl vinyl ether	0.159	0.129	18.9	82	0.00
56 M	Bromoform	0.305	0.246	19.3	84	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
58 M	4-Methyl-2-Pentanone	0.569	0.500	12.1	86	0.00
59 M	2-Hexanone	0.416	0.358	13.9	84	0.00
60 S	4-Bromofluorobenzene	0.445	0.442	0.7	96	0.00
61 M	Tetrachloroethene	0.365	0.315	13.7	82	0.00
62 M	Toluene	1.819	1.563	14.1	82	0.00
63 S	Toluene-d8	1.362	1.396	-2.5	96	0.00
64 M	Chlorobenzene	1.122	0.956	14.8	84	0.00
65	1,1,1,2-Tetrachloroethane	0.412	0.363	11.9	85	0.00
66 M	Ethyl Benzene	1.977	1.652	16.4	83	0.00
67 M	m/p-Xylenes	0.769	0.634	17.6	80	0.00
68 M	o-Xylene	0.764	0.640	16.2	82	0.00
69 M	Styrene	1.203	0.985	18.1	83	0.00
70	Isopropylbenzene	1.918	1.608	16.2	82	0.00
71 M	1,1,2,2-Tetrachloroethane	0.624	0.555	11.1	84	0.00
72	1,2,3-Trichloropropane	0.536	0.492	8.2	86	0.00
73	Bromobenzene	0.468	0.389	16.9	84	0.00
74	n-propylbenzene	2.094	1.684	19.6	81	0.00
75	2-Chlorotoluene	1.308	1.100	15.9	83	0.00
76	1,3,5-Trimethylbenzene	1.539	1.289	16.2	82	0.00
77	t-1,4-Dichloro-2-butene	0.188	0.142	24.5	83	0.00
78	4-Chlorotoluene	1.222	1.002	18.0	83	0.00
79	tert-butylbenzene	1.389	1.166	16.1	83	0.00
80	1,2,4-Trimethylbenzene	1.493	1.248	16.4	83	0.00
81	sec-Butylbenzene	1.866	1.528	18.1	82	0.00
82	p-Isopropyltoluene	1.516	1.216	19.8	82	0.00
83 M	1,3-Dichlorobenzene	0.766	0.627	18.1	84	0.00
84 M	1,4-Dichlorobenzene	0.731	0.600	17.9	85	0.00
85	n-Butylbenzene	1.126	0.847	24.8	83	0.00
86 T	Hexachloroethane	0.283	0.231	18.4	83	0.00
87 M	1,2-Dichlorobenzene	0.740	0.619	16.4	82	0.00
88	1,2-Dibromo-3-Chloropropane	0.111	0.094	15.3	88	0.00
89	1,2,4-Trichlorobenzene	0.342	0.272	20.5	88	0.00
90	Hexachlorobutadiene	0.203	0.181	10.8	89	0.00
91 M	Naphthalene	0.951	0.709	25.4	90	0.00
92	1,2,3-Trichlorobenzene	0.336	0.269	19.9	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0