

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032624\
 Data File : VN081578.D
 Acq On : 26 Mar 2024 14:13
 Operator : JC\MD
 Sample : P1838-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-SP302-20240321

Quant Time: Mar 27 04:37:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	411342	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	721063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	631801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	249332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	261646	49.878	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.760%
35) Dibromofluoromethane	8.171	113	226538	49.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.571	98	847700	51.532	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.847	95	260963	45.022	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.040%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	15654	2.987	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032624\
Data File : VN081578.D
Acq On : 26 Mar 2024 14:13
Operator : JC\MD
Sample : P1838-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137-SP302-20240321

Quant Time: Mar 27 04:37:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

