

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032624\
 Data File : VN081573.D
 Acq On : 26 Mar 2024 11:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 04:34:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.673	0.673	0.0	92	0.00
3 P	Chloromethane	0.670	0.648	3.3	91	0.00
4 C	Vinyl Chloride	0.752	0.772	-2.7#	96	0.00
5 T	Bromomethane	0.532	0.548	-3.0	96	0.01
6 T	Chloroethane	0.517	0.528	-2.1	95	0.00
7 T	Trichlorofluoromethane	1.121	1.223	-9.1	99	0.00
8 T	Diethyl Ether	0.371	0.412	-11.1	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.637	0.714	-12.1	103	0.00
10 T	Methyl Iodide	0.634	0.749	-18.1	97	0.00
11 T	Tert butyl alcohol	0.111	0.110	0.9	86	0.00
12 CM	1,1-Dichloroethene	0.590	0.648	-9.8#	99	0.00
13 T	Acrolein	0.100	0.125	-25.0	101	0.00
14 T	Allyl chloride	0.739	0.873	-18.1	105	0.00
15 T	Acrylonitrile	0.297	0.287	3.4	87	0.00
16 T	Acetone	0.206	0.175	15.0	80	0.00
17 T	Carbon Disulfide	1.693	1.655	2.2	92	0.00
18 T	Methyl Acetate	0.635	0.681	-7.2	94	0.00
19 T	Methyl tert-butyl Ether	1.847	2.128	-15.2	99	0.00
20 T	Methylene Chloride	0.675	0.717	-6.2	100	0.00
21 T	trans-1,2-Dichloroethene	0.658	0.716	-8.8	101	0.00
22 T	Diisopropyl ether	1.718	2.014	-17.2	100	0.00
23 T	Vinyl Acetate	1.368	1.512	-10.5	92	0.00
24 P	1,1-Dichloroethane	1.151	1.271	-10.4	101	0.00
25 T	2-Butanone	0.353	0.318	9.9	79	0.00
26 T	2,2-Dichloropropane	0.998	1.208	-21.0	103	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.826	-11.0	100	0.00
28 T	Bromochloromethane	0.469	0.393	16.2	75	0.00
29 T	Tetrahydrofuran	0.238	0.224	5.9	81	0.00
30 C	Chloroform	1.232	1.395	-13.2#	102	0.00
31 T	Cyclohexane	1.038	1.027	1.1	93	0.00
32 T	1,1,1-Trichloroethane	1.118	1.247	-11.5	101	0.00
33 S	1,2-Dichloroethane-d4	0.638	0.624	2.2	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.316	0.338	-7.0	89	0.00
36 T	1,1-Dichloropropene	0.531	0.608	-14.5	100	0.00
37 T	Ethyl Acetate	0.455	0.436	4.2	84	0.00
38 T	Carbon Tetrachloride	0.582	0.666	-14.4	100	0.00
39 T	Methylcyclohexane	0.593	0.716	-20.7	99	0.00
40 TM	Benzene	1.553	1.772	-14.1	98	0.00
41 T	Methacrylonitrile	0.240	0.239	0.4	89	0.00
42 TM	1,2-Dichloroethane	0.546	0.602	-10.3	98	0.00
43 T	Isopropyl Acetate	0.754	0.769	-2.0	87	0.00
44 TM	Trichloroethene	0.437	0.493	-12.8	101	0.00
45 C	1,2-Dichloropropane	0.376	0.439	-16.8#	100	0.00
46 T	Dibromomethane	0.286	0.318	-11.2	100	0.00
47 T	Bromodichloromethane	0.559	0.645	-15.4	99	0.00
48 T	Methyl methacrylate	0.348	0.360	-3.4	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032624\
 Data File : VN081573.D
 Acq On : 26 Mar 2024 11:53
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 27 04:34:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	72	0.00
50 S	Toluene-d8	1.141	1.189	-4.2	86	0.00
51 T	4-Methyl-2-Pentanone	0.460	0.447	2.8	82	0.00
52 CM	Toluene	0.987	1.138	-15.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.603	0.703	-16.6	98	0.00
54 T	cis-1,3-Dichloropropene	0.647	0.764	-18.1	101	0.00
55 T	1,1,2-Trichloroethane	0.397	0.440	-10.8	99	0.00
56 T	Ethyl methacrylate	0.529	0.625	-18.1	94	0.00
57 T	1,3-Dichloropropane	0.651	0.754	-15.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.268	-7.2	77	0.00
59 T	2-Hexanone	0.340	0.326	4.1	78	0.00
60 T	Dibromochloromethane	0.437	0.497	-13.7	99	0.00
61 T	1,2-Dibromoethane	0.405	0.448	-10.6	94	0.00
62 S	4-Bromofluorobenzene	0.402	0.415	-3.2	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.478	0.562	-17.6	102	0.00
65 PM	Chlorobenzene	1.182	1.366	-15.6	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.438	0.525	-19.9	99	0.00
67 C	Ethyl Benzene	2.053	2.489	-21.2#	97	0.00
68 T	m/p-Xylenes	0.770	0.961	-24.8	99	0.00
69 T	o-Xylene	0.756	0.918	-21.4	98	0.00
70 T	Styrene	1.219	1.542	-26.5#	98	0.00
71 P	Bromoform	0.329	0.358	-8.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	4.262	5.046	-18.4	97	0.00
74 T	N-amyl acetate	1.542	1.623	-5.3	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.253	1.247	0.5	88	0.00
76 T	1,2,3-Trichloropropane	1.196	1.122	6.2	83	0.00
77 T	Bromobenzene	1.053	1.165	-10.6	97	0.00
78 T	n-propylbenzene	5.033	5.980	-18.8	96	0.00
79 T	2-Chlorotoluene	3.012	3.427	-13.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.520	4.261	-21.1	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.455	-2.7	83	0.00
82 T	4-Chlorotoluene	2.970	3.508	-18.1	98	0.00
83 T	tert-Butylbenzene	3.036	3.625	-19.4	98	0.00
84 T	1,2,4-Trimethylbenzene	3.538	4.239	-19.8	97	0.00
85 T	sec-Butylbenzene	4.340	5.235	-20.6	98	0.00
86 T	p-Isopropyltoluene	3.547	4.405	-24.2	99	0.00
87 T	1,3-Dichlorobenzene	1.934	2.217	-14.6	101	0.00
88 T	1,4-Dichlorobenzene	2.003	2.208	-10.2	100	0.00
89 T	n-Butylbenzene	3.146	3.877	-23.2	98	0.00
90 T	Hexachloroethane	0.672	0.767	-14.1	98	0.00
91 T	1,2-Dichlorobenzene	1.862	2.164	-16.2	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.267	0.267	0.0	86	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.331	-23.6	104	0.00
94 T	Hexachlorobutadiene	0.472	0.537	-13.8	99	0.00
95 T	Naphthalene	3.488	4.158	-19.2	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032624\
Data File : VN081573.D
Acq On : 26 Mar 2024 11:53
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 27 04:34:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.083	1.312	-21.1	104	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6