

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
 Data File : VN086138.D
 Acq On : 26 Mar 2025 20:00
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Mar 27 01:45:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	298261	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	278061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	140375	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135243	58.226	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 116.460%	
35) Dibromofluoromethane	8.165	113	110596	53.125	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.240%	
50) Toluene-d8	10.565	98	381848	50.539	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.080%	
62) 4-Bromofluorobenzene	12.847	95	140691	52.213	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.420%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	82011	36.494	ug/l	96
3) Chloromethane	2.359	50	108979	55.317	ug/l	99
4) Vinyl Chloride	2.512	62	103084	51.175	ug/l	99
5) Bromomethane	2.947	94	73999	53.931	ug/l	100
6) Chloroethane	3.112	64	71667	56.397	ug/l	93
7) Trichlorofluoromethane	3.495	101	145534	40.191	ug/l	100
8) Diethyl Ether	3.959	74	70153	63.285	ug/l	92
9) 1,1,2-Trichlorotrifluo...	4.371	101	84274	43.790	ug/l	97
10) Methyl Iodide	4.583	142	138620	55.279	ug/l	99
11) Tert butyl alcohol	5.518	59	94924	288.650	ug/l	97
12) 1,1-Dichloroethene	4.336	96	84805	48.801	ug/l	98
13) Acrolein	4.183	56	61775	176.122	ug/l	97
14) Allyl chloride	5.024	41	126162	58.193	ug/l	94
15) Acrylonitrile	5.718	53	294081	338.606	ug/l	98
16) Acetone	4.424	43	211763	298.421	ug/l	99
17) Carbon Disulfide	4.706	76	238384	41.712	ug/l	98
18) Methyl Acetate	5.018	43	132558	75.502	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	369462	64.881	ug/l	97
20) Methylene Chloride	5.271	84	123830	59.651	ug/l	94
21) trans-1,2-Dichloroethene	5.788	96	101739	53.571	ug/l	98
22) Diisopropyl ether	6.671	45	322564	69.820	ug/l	98
23) Vinyl Acetate	6.600	43	1205462	344.201	ug/l	97
24) 1,1-Dichloroethane	6.565	63	199407	58.696	ug/l	99
25) 2-Butanone	7.482	43	338700	330.785	ug/l	93
26) 2,2-Dichloropropane	7.488	77	144911	44.305	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	128830	59.696	ug/l	98
28) Bromochloromethane	7.812	49	93030	64.633	ug/l	94
29) Tetrahydrofuran	7.835	42	231334	365.956	ug/l	94
30) Chloroform	7.965	83	213996	56.988	ug/l	99
31) Cyclohexane	8.253	56	123098	42.534	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	169265	47.619	ug/l	99
36) 1,1-Dichloropropene	8.365	75	122330	44.626	ug/l	99
37) Ethyl Acetate	7.553	43	135938	58.781	ug/l	98
38) Carbon Tetrachloride	8.359	117	149162	41.385	ug/l	97
39) Methylcyclohexane	9.600	83	108557	39.915	ug/l	92
40) Benzene	8.606	78	451598	52.478	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
 Data File : VN086138.D
 Acq On : 26 Mar 2025 20:00
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Mar 27 01:45:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	73006	66.111	ug/l	96
42) 1,2-Dichloroethane	8.665	62	158969	53.016	ug/l	99
43) Isopropyl Acetate	8.688	43	248862	49.660	ug/l	94
44) Trichloroethene	9.347	130	97307	43.629	ug/l	95
45) 1,2-Dichloropropane	9.618	63	112520	57.488	ug/l	95
46) Dibromomethane	9.706	93	86713	56.335	ug/l	98
47) Bromodichloromethane	9.888	83	175292	54.490	ug/l	97
48) Methyl methacrylate	9.676	41	105777	63.964	ug/l	98
49) 1,4-Dioxane	9.694	88	47420	1213.018	ug/l	91
51) 4-Methyl-2-Pentanone	10.441	43	729178	329.927	ug/l	98
52) Toluene	10.629	92	277758	52.590	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	173326	55.478	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	187717	56.949	ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	118547	58.808	ug/l	96
56) Ethyl methacrylate	10.871	69	183825	56.881	ug/l	98
57) 1,3-Dichloropropane	11.159	76	199137	58.161	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	414605	324.376	ug/l	96
59) 2-Hexanone	11.194	43	534086	331.719	ug/l	98
60) Dibromochloromethane	11.359	129	143925	55.329	ug/l	98
61) 1,2-Dibromoethane	11.465	107	117007	56.539	ug/l	97
64) Tetrachloroethene	11.100	164	86768	39.118	ug/l	98
65) Chlorobenzene	11.888	112	302126	47.483	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.959	131	114808	48.539	ug/l	95
67) Ethyl Benzene	11.965	91	481325	47.290	ug/l	99
68) m/p-Xylenes	12.070	106	389180	97.284	ug/l	98
69) o-Xylene	12.394	106	189393	51.583	ug/l	98
70) Styrene	12.412	104	334569	53.349	ug/l	98
71) Bromoform	12.576	173	103058	52.121	ug/l #	99
73) Isopropylbenzene	12.694	105	437333	45.622	ug/l	99
74) N-amyl acetate	12.494	43	195624	60.996	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	172050	53.051	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	222999	70.957	ug/l	68
77) Bromobenzene	12.976	156	126286	47.246	ug/l	95
78) n-propylbenzene	13.035	91	505397	46.560	ug/l	99
79) 2-Chlorotoluene	13.123	91	341837	47.412	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	382694	47.285	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	65016	54.695	ug/l	96
82) 4-Chlorotoluene	13.217	91	346312	46.709	ug/l	98
83) tert-Butylbenzene	13.435	119	301237	43.519	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	401306	49.161	ug/l	99
85) sec-Butylbenzene	13.611	105	414364	45.888	ug/l	99
86) p-Isopropyltoluene	13.729	119	351263	45.953	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	220509	45.768	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	218845	43.232	ug/l	99
89) n-Butylbenzene	14.053	91	269430	43.284	ug/l	99
90) Hexachloroethane	14.329	117	71052	40.284	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	226649	47.321	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	33601	51.904	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	102605	42.605	ug/l	99
94) Hexachlorobutadiene	15.500	225	47400	33.925	ug/l	100
95) Naphthalene	15.641	128	346355	49.667	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	103041	45.091	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
Data File : VN086138.D
Acq On : 26 Mar 2025 20:00
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 27 01:45:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
 Data File : VN086138.D
 Acq On : 26 Mar 2025 20:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 27 01:45:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 169623
Ion Ratio Lower Upper
168 100
99 60.7 49.4 74.2



#2
Dichlorodifluoromethane
Concen: 36.494 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 85 Resp: 82011
Ion Ratio Lower Upper
85 100
87 30.9 16.5 49.5









#3
Chloromethane
Concen: 55.317 ug/l
RT: 2.359 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 108979
Ion Ratio Lower Upper
50 100
52 31.9 26.0 39.0



#4
Vinyl Chloride
Concen: 51.175 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 62 Resp: 103084
Ion Ratio Lower Upper
62 100
64 31.0 25.2 37.8









#5
Bromomethane
Concen: 53.931 ug/l
RT: 2.947 min Scan# 168
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 73999
Ion Ratio Lower Upper
94 100
96 90.8 72.8 109.2



#6
Chloroethane
Concen: 56.397 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 64 Resp: 71667
Ion Ratio Lower Upper
64 100
66 31.3 28.3 42.5









#7
Trichlorofluoromethane
Concen: 40.191 ug/l
RT: 3.495 min Scan# 261
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 145534
Ion Ratio Lower Upper
101 100
103 64.5 51.7 77.5



#8
Diethyl Ether
Concen: 63.285 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 74 Resp: 70153
Ion Ratio Lower Upper
74 100
45 82.6 38.0 113.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 43.790 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 84274
Ion Ratio Lower Upper
101 100
85 44.4 36.5 54.7
151 76.6 63.4 95.0



#10
Methyl Iodide
Concen: 55.279 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:142 Resp: 138620
Ion Ratio Lower Upper
142 100
127 47.6 37.5 56.3
141 14.3 11.0 16.6









#11
Tert butyl alcohol
Concen: 288.650 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 94924
Ion Ratio Lower Upper
59 100
57 11.0 7.9 11.9



#12
1,1-Dichloroethene
Concen: 48.801 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 96 Resp: 84805
Ion Ratio Lower Upper
96 100
61 149.5 120.5 180.7
98 63.8 47.4 71.0









#13
Acrolein
Concen: 176.122 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 56 Resp: 61775
Ion Ratio Lower Upper
56 100
55 71.5 55.1 82.7



#14
Allyl chloride
Concen: 58.193 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 41 Resp: 126162
Ion Ratio Lower Upper
41 100
39 63.3 55.5 83.3
76 46.0 39.1 58.7









#15
Acrylonitrile
Concen: 338.606 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 294081
Ion Ratio Lower Upper
53 100
52 81.2 63.9 95.9
51 34.9 27.1 40.7



#16
Acetone
Concen: 298.421 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 43 Resp: 211763
Ion Ratio Lower Upper
43 100
58 37.5 29.4 44.2









#17
Carbon Disulfide
Concen: 41.712 ug/l
RT: 4.706 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 238384
Ion Ratio Lower Upper
76 100
78 8.5 7.4 11.2



#18
Methyl Acetate
Concen: 75.502 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 43 Resp: 132558
Ion Ratio Lower Upper
43 100
74 28.6 23.9 35.9









#19
Methyl tert-butyl Ether
Concen: 64.881 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 369462
Ion Ratio Lower Upper
73 100
57 21.5 16.1 24.1



#20
Methylene Chloride
Concen: 59.651 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 84 Resp: 123830
Ion Ratio Lower Upper
84 100
49 113.1 83.3 124.9
51 33.8 24.8 37.2
86 63.3 48.7 73.1









#21
trans-1,2-Dichloroethene
Concen: 53.571 ug/l
RT: 5.788 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 101739
Ion Ratio Lower Upper
96 100
61 134.4 106.6 160.0
98 65.2 50.0 75.0



#22
Diisopropyl ether
Concen: 69.820 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 45 Resp: 322564
Ion Ratio Lower Upper
45 100
43 52.9 41.2 61.8
87 34.1 28.5 42.7
59 13.3 11.2 16.8









#23
 Vinyl Acetate
 Concen: 344.201 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1205462
 Ion Ratio Lower Upper
 43 100
 86 12.7 11.2 16.8



#24
 1,1-Dichloroethane
 Concen: 58.696 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00

Tgt Ion: 63 Resp: 199407
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.6 10.9
 100 4.4 2.5 7.5









#25
2-Butanone
Concen: 330.785 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 338700
Ion Ratio Lower Upper
43 100
72 30.3 27.4 41.0



#26
2,2-Dichloropropane
Concen: 44.305 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 77 Resp: 144911
Ion Ratio Lower Upper
77 100
97 22.8 11.2 33.5









#27
 cis-1,2-Dichloroethene
 Concen: 59.696 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 128830
 Ion Ratio Lower Upper
 96 100
 61 134.7 0.0 272.8
 98 64.5 0.0 124.8



#28
 Bromochloromethane
 Concen: 64.633 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00

Tgt Ion: 49 Resp: 93030
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.2
 130 86.8 74.2 111.4





#29

Tetrahydrofuran

Concen: 365.956 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 231334

Ion Ratio Lower Upper

42 100

72 52.3 45.0 67.4

71 48.3 41.9 62.9



#30

Chloroform

Concen: 56.988 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN086138.D




Acq: 26 Mar 2025 20:00


Tgt Ion: 83 Resp: 213996

Ion Ratio Lower Upper

83 100

85 63.5 51.5 77.3





#31

Cyclohexane

Concen: 42.534 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 56 Resp: 123098

Ion Ratio Lower Upper

56 100

69 32.4 28.0 42.0

84 93.1 75.0 112.6



#32

1,1,1-Trichloroethane

Concen: 47.619 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion: 97 Resp: 169265


Ion Ratio Lower Upper

97 100

99 65.1 52.8 79.2

61 42.4 33.6 50.4





#33

1,2-Dichloroethane-d4

Concen: 58.226 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 135243

Ion Ratio Lower Upper

65 100

67 52.6 0.0 102.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion: 114 Resp: 298261


Ion Ratio Lower Upper

114 100

63 20.6 0.0 39.6

88 16.3 0.0 32.6





#35

Dibromofluoromethane

Concen: 53.125 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 110596

Ion Ratio Lower Upper

113 100

111 101.5 81.8 122.8

192 19.6 15.9 23.9



#36

1,1-Dichloropropene

Concen: 44.626 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion: 75 Resp: 122330


Ion Ratio Lower Upper

75 100

110 34.8 17.8 53.3

77 31.2 25.0 37.4









#37
Ethyl Acetate
Concen: 58.781 ug/l
RT: 7.553 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 135938
Ion Ratio Lower Upper
43 100
61 15.9 13.0 19.4
70 12.8 11.0 16.6



#38
Carbon Tetrachloride
Concen: 41.385 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 117 Resp: 149162
Ion Ratio Lower Upper
117 100
119 92.5 77.0 115.4
121 32.1 25.6 38.4









#39
Methylcyclohexane
Concen: 39.915 ug/l
RT: 9.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 108557
Ion Ratio Lower Upper
83 100
55 76.3 54.2 81.2
98 49.4 37.1 55.7



#40
Benzene
Concen: 52.478 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 78 Resp: 451598
Ion Ratio Lower Upper
78 100
77 23.9 18.9 28.3









#41
Methacrylonitrile
Concen: 66.111 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:	41	Resp:	73006
Ion	Ratio	Lower	Upper
41	100		
39	54.6	46.2	69.2
67	89.8	75.9	113.9
52	36.1	29.2	43.8



#42
1,2-Dichloroethane
Concen: 53.016 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:	62	Resp:	158969
Ion	Ratio	Lower	Upper
62	100		
98	9.8	0.0	18.6









#43
Isopropyl Acetate
Concen: 49.660 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 248862
Ion Ratio Lower Upper
43 100
61 27.2 25.0 37.6
87 14.9 13.0 19.4



#44
Trichloroethene
Concen: 43.629 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 130 Resp: 97307
Ion Ratio Lower Upper
130 100
95 100.4 0.0 191.6





#45

1,2-Dichloropropane

Concen: 57.488 ug/l

RT: 9.618 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 112520

Ion Ratio Lower Upper

63 100

65 33.6 24.7 37.1



#46

Dibromomethane

Concen: 56.335 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion: 93 Resp: 86713


Ion Ratio Lower Upper

93 100

95 82.3 66.2 99.2

174 97.4 80.1 120.1









#47
Bromodichloromethane
Concen: 54.490 ug/l
RT: 9.888 min Scan# 1348
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 175292
Ion Ratio Lower Upper
83 100
85 65.0 50.2 75.2
127 8.6 7.1 10.7



#48
Methyl methacrylate
Concen: 63.964 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 41 Resp: 105777
Ion Ratio Lower Upper
41 100
69 103.3 83.1 124.7
39 53.6 45.6 68.4





#49

1,4-Dioxane

Concen: 1213.018 ug/l

RT: 9.694 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 47420

Ion Ratio Lower Upper

88 100

43 22.5 21.7 32.5

58 70.3 50.4 75.6



#50

Toluene-d8

Concen: 50.539 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN086138.D




Acq: 26 Mar 2025 20:00


Tgt Ion: 98 Resp: 381848

Ion Ratio Lower Upper

98 100

100 65.0 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 329.927 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 729178
Ion Ratio Lower Upper
43 100
58 45.2 37.4 56.0



#52
Toluene
Concen: 52.590 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 92 Resp: 277758
Ion Ratio Lower Upper
92 100
91 169.5 136.2 204.4









#53
 t-1,3-Dichloropropene
 Concen: 55.478 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 173326
 Ion Ratio Lower Upper
 75 100
 77 30.9 25.3 37.9



#54
 cis-1,3-Dichloropropene
 Concen: 56.949 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00

Tgt Ion: 75 Resp: 187717
 Ion Ratio Lower Upper
 75 100
 77 30.4 25.3 37.9
 39 34.7 29.6 44.4





#55

1,1,2-Trichloroethane

Concen: 58.808 ug/l

RT: 11.012 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 118547





Ion Ratio Lower Upper

97 100

83 84.1 71.2 106.8

85 54.8 45.4 68.0

99 63.8 49.5 74.3



#56

Ethyl methacrylate

Concen: 56.881 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion: 69 Resp: 183825


Ion Ratio Lower Upper

69 100

41 53.6 41.6 62.4

39 26.9 21.8 32.8









#57
1,3-Dichloropropane
Concen: 58.161 ug/l
RT: 11.159 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 199137
Ion Ratio Lower Upper
76 100
78 33.2 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 324.376 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 63 Resp: 414605
Ion Ratio Lower Upper
63 100
106 29.3 25.0 37.6









#59
2-Hexanone
Concen: 331.719 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 534086
Ion Ratio Lower Upper
43 100
58 63.1 32.4 97.2



#60
Dibromochloromethane
Concen: 55.329 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:129 Resp: 143925
Ion Ratio Lower Upper
129 100
127 77.5 37.8 113.4









#61
1,2-Dibromoethane
Concen: 56.539 ug/l
RT: 11.465 min Scan# 117007
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 117007
Ion Ratio Lower Upper
107 100
109 94.4 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 52.213 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 95 Resp: 140691
Ion Ratio Lower Upper
95 100
174 77.6 0.0 156.8
176 73.8 0.0 152.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 278061
Ion Ratio Lower Upper
117 100
82 53.3 46.7 70.1
119 32.7 26.5 39.7



#64
Tetrachloroethene
Concen: 39.118 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:164 Resp: 86768
Ion Ratio Lower Upper
164 100
166 128.9 103.7 155.5
129 97.0 77.1 115.7
131 95.1 73.3 109.9









#65
Chlorobenzene
Concen: 47.483 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:112 Resp: 302126
Ion Ratio Lower Upper
112 100
114 30.8 26.3 39.5



#66
1,1,1,2-Tetrachloroethane
Concen: 48.539 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:131 Resp: 114808
Ion Ratio Lower Upper
131 100
133 98.9 47.0 141.0
119 65.5 31.4 94.0









#67
Ethyl Benzene
Concen: 47.290 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 481325
Ion Ratio Lower Upper
91 100
106 31.1 24.5 36.7



#68
m/p-Xylenes
Concen: 97.284 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 106 Resp: 389180
Ion Ratio Lower Upper
106 100
91 202.3 163.9 245.9









#69
o-Xylene
Concen: 51.583 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 189393
Ion Ratio Lower Upper
106 100
91 215.3 109.1 327.2



#70
Styrene
Concen: 53.349 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:104 Resp: 334569
Ion Ratio Lower Upper
104 100
78 49.6 41.4 62.2
103 53.7 42.9 64.3





#71

Bromoform

Concen: 52.121 ug/l

RT: 12.576 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 103058

Ion Ratio Lower Upper

173 100

175 48.1 24.3 73.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion:152 Resp: 140375


Ion Ratio Lower Upper

152 100

115 61.9 31.1 93.2

150 173.9 0.0 359.6









#73
Isopropylbenzene
Concen: 45.622 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 437333
Ion Ratio Lower Upper
105 100
120 26.6 13.0 39.0



#74
N-ethyl acetate
Concen: 60.996 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 43 Resp: 195624
Ion Ratio Lower Upper
43 100
70 53.6 43.0 64.4
55 31.0 23.8 35.8
61 28.7 23.4 35.0





#75

1,1,2,2-Tetrachloroethane

Concen: 53.051 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 172050

Ion Ratio Lower Upper

83 100

131 10.4 5.5 16.5

85 65.6 32.6 97.8



#76

1,2,3-Trichloropropane

Concen: 70.957 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN086138.D




Acq: 26 Mar 2025 20:00


Tgt Ion: 75 Resp: 222999

Ion Ratio Lower Upper

75 100

77 124.1 83.7 251.0









#77
Bromobenzene
Concen: 47.246 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 126286
Ion Ratio Lower Upper
156 100
77 219.1 104.1 312.3
158 98.6 49.0 147.0



#78
n-propylbenzene
Concen: 46.560 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 91 Resp: 505397
Ion Ratio Lower Upper
91 100
120 22.7 11.7 35.0









#79
2-Chlorotoluene
Concen: 47.412 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 341837
Ion Ratio Lower Upper
91 100
126 32.8 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 47.285 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 105 Resp: 382694
Ion Ratio Lower Upper
105 100
120 48.3 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 54.695 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 65016
Ion Ratio Lower Upper
75 100
53 86.6 66.0 99.0
89 44.0 36.9 55.3



#82
4-Chlorotoluene
Concen: 46.709 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 91 Resp: 346312
Ion Ratio Lower Upper
91 100
126 33.2 16.1 48.2









#83
tert-Butylbenzene
Concen: 43.519 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 301237
Ion Ratio Lower Upper
119 100
91 67.7 34.3 102.9
134 27.4 13.5 40.5



#84
1,2,4-Trimethylbenzene
Concen: 49.161 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion:105 Resp: 401306
Ion Ratio Lower Upper
105 100
120 44.5 21.9 65.8









#85
 sec-Butylbenzene
 Concen: 45.888 ug/l
 RT: 13.611 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 414364
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 45.953 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00

Tgt Ion:119 Resp: 351263
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.6 37.8
 91 24.3 12.3 36.8









#87
 1,3-Dichlorobenzene
 Concen: 45.768 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 220509
 Ion Ratio Lower Upper
 146 100
 111 42.6 21.2 63.6
 148 64.7 31.8 95.4



#88
 1,4-Dichlorobenzene
 Concen: 43.232 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN086138.D
 Acq: 26 Mar 2025 20:00

Tgt Ion:146 Resp: 218845
 Ion Ratio Lower Upper
 146 100
 111 41.6 20.4 61.3
 148 65.2 32.1 96.5





#89

n-Butylbenzene

Concen: 43.284 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 269430

Ion Ratio Lower Upper

91 100

92 52.6 26.9 80.6

134 26.0 13.0 39.0



#90

Hexachloroethane

Concen: 40.284 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.000 min

Lab File: VN086138.D




Acq: 26 Mar 2025 20:00


Tgt Ion: 117 Resp: 71052

Ion Ratio Lower Upper

117 100

201 76.1 39.1 117.3









#91
1,2-Dichlorobenzene
Concen: 47.321 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 226649
Ion Ratio Lower Upper
146 100
111 43.4 22.3 66.8
148 63.2 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.904 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086138.D
Acq: 26 Mar 2025 20:00

Tgt Ion: 75 Resp: 33601
Ion Ratio Lower Upper
75 100
155 81.5 41.0 123.0
157 103.2 53.7 161.1





#93

1,2,4-Trichlorobenzene

Concen: 42.605 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 102605

Ion Ratio Lower Upper

180 100

182 94.6 47.0 141.0

145 31.9 15.5 46.5



#94

Hexachlorobutadiene

Concen: 33.925 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00




Tgt Ion:225 Resp: 47400


Ion Ratio Lower Upper

225 100

223 62.9 31.3 93.9

227 64.6 32.1 96.3





#95

Naphthalene

Concen: 49.667 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 346355

Ion Ratio Lower Upper

128 100

127 13.0 10.2 15.4

129 10.7 9.0 13.6



#96

1,2,3-Trichlorobenzene

Concen: 45.091 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN086138.D

Acq: 26 Mar 2025 20:00

Tgt Ion:180 Resp: 103041

Ion Ratio Lower Upper

180 100

182 97.5 47.5 142.5

145 34.3 17.7 53.1

