

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
 Data File : VN086130.D
 Acq On : 26 Mar 2025 16:48
 Operator : JC\MD
 Sample : Q1636-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Mar 27 01:41:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129785	53.430	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.860%	
35) Dibromofluoromethane	8.165	113	114159	50.467	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.940%	
50) Toluene-d8	10.565	98	417529	50.858	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.720%	
62) 4-Bromofluorobenzene	12.847	95	137125	46.835	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.660%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	101153	136.305	ug/l	98
20) Methylene Chloride	5.271	84	5903	2.719	ug/l	92
25) 2-Butanone	7.489	43	13657	12.754	ug/l	94
43) Isopropyl Acetate	8.688	43	29709	5.456	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
Data File : VN086130.D
Acq On : 26 Mar 2025 16:48
Operator : JC\MD
Sample : Q1636-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Mar 27 01:41:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

