

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
 Data File : VN086133.D
 Acq On : 26 Mar 2025 18:00
 Operator : JC\MD
 Sample : Q1641-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NP-WS-003

Quant Time: Mar 27 01:43:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125717	54.405	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.800%		
35) Dibromofluoromethane	8.165	113	109372	50.142	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.280%		
50) Toluene-d8	10.565	98	375240	47.400	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.800%		
62) 4-Bromofluorobenzene	12.847	95	128545	45.531	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 91.060%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.518	59	47221	144.334	ug/l	99
16) Acetone	4.430	43	6471	9.166	ug/l	97
20) Methylene Chloride	5.277	84	11751	5.690	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032625\
Data File : VN086133.D
Acq On : 26 Mar 2025 18:00
Operator : JC\MD
Sample : Q1641-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NP-WS-003

Quant Time: Mar 27 01:43:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

