

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032719\
 Data File : VN054718.D
 Acq On : 27 Mar 2019 13:31
 Operator : JC/SP
 Sample : K2086-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-13

Quant Time: Mar 28 07:07:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	529300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	877711	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	727416	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	254669	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	357421	54.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.58%	
35) Dibromofluoromethane	7.59	113	283927	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
50) Toluene-d8	10.09	98	1045630	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
62) 4-Bromofluorobenzene	12.40	95	313796	43.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.76%	
Target Compounds						
16) Acetone	3.82	43	14046	7.096	ug/l #	71
20) Methylene Chloride	4.55	84	18130	3.107	ug/l #	81
43) Isopropyl Acetate	8.17	43	13573	1.377	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032719\
Data File : VN054718.D
Acq On : 27 Mar 2019 13:31
Operator : JC/SP
Sample : K2086-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-13

Quant Time: Mar 28 07:07:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

