

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032719\
 Data File : VN054720.D
 Acq On : 27 Mar 2019 14:24
 Operator : JC/SP
 Sample : K2086-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-14

Quant Time: Mar 28 07:11:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	546795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	887596	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	728851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	253554	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	361626	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
35) Dibromofluoromethane	7.59	113	286609	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	10.09	98	1048764	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
62) 4-Bromofluorobenzene	12.40	95	317589	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
Target Compounds						
16) Acetone	3.82	43	11591	5.669	ug/l #	74
20) Methylene Chloride	4.55	84	18583	3.083	ug/l #	87
43) Isopropyl Acetate	8.17	43	13764	1.381	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032719\
Data File : VN054720.D
Acq On : 27 Mar 2019 14:24
Operator : JC/SP
Sample : K2086-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-14

Quant Time: Mar 28 07:11:46 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

