

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032719\
 Data File : VN054723.D
 Acq On : 27 Mar 2019 15:43
 Operator : JC/SP
 Sample : K2086-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-18

Quant Time: Mar 28 07:17:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	541086	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	862692	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	742371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	267126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	351219	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	7.59	113	273194	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	10.09	98	1036171	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	12.40	95	320789	45.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.24%	
Target Compounds						
16) Acetone	3.83	43	17014	8.408	ug/l #	84
20) Methylene Chloride	4.54	84	20395	3.419	ug/l #	78
43) Isopropyl Acetate	8.17	43	13232	1.366	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032719\
Data File : VN054723.D
Acq On : 27 Mar 2019 15:43
Operator : JC/SP
Sample : K2086-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-18

Quant Time: Mar 28 07:17:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

