

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032719\
 Data File : VN054730.D
 Acq On : 27 Mar 2019 18:47
 Operator : JC/SP
 Sample : K2086-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-12

Quant Time: Mar 28 07:30:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	508851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	841714	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	724881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	261105	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	337931	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
35) Dibromofluoromethane	7.59	113	271897	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.09	98	1022218	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.40	95	312188	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
Target Compounds						
16) Acetone	3.82	43	12577	6.609	ug/l #	75
20) Methylene Chloride	4.55	84	17690	3.154	ug/l #	79
43) Isopropyl Acetate	8.17	43	14536	1.538	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032719\
Data File : VN054730.D
Acq On : 27 Mar 2019 18:47
Operator : JC/SP
Sample : K2086-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-12

Quant Time: Mar 28 07:30:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

