

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032720\  
 Data File : VN060738.D  
 Acq On : 27 Mar 2020 13:45  
 Operator : JC/MD  
 Sample : PB127814ZHE#03  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB127814ZHE#03

Quant Time: Mar 30 00:43:50 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N031820W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 18 08:49:09 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.64  | 168  | 191099   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.57  | 114  | 317966   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.40 | 117  | 299755   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.34 | 152  | 136745   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 8.00  | 65   | 123175   | 46.95  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 93.90%  |          |
| 35) Dibromofluoromethane    | 7.56  | 113  | 93771    | 48.77  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 97.54%  |          |
| 50) Toluene-d8              | 10.08 | 98   | 391222   | 49.82  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.64%  |          |
| 62) 4-Bromofluorobenzene    | 12.40 | 95   | 147316   | 51.19  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.38% |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 3.78  | 43   | 5825     | 6.269  | ug/l    | 91       |
| 20) Methylene Chloride      | 4.52  | 84   | 4374     | 1.913  | ug/l    | 95       |
| 43) Isopropyl Acetate       | 8.14  | 43   | 59091    | 10.770 | ug/l #  | 91       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032720\  
Data File : VN060738.D  
Acq On : 27 Mar 2020 13:45  
Operator : JC/MD  
Sample : PB127814ZHE#03  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB127814ZHE#03

Quant Time: Mar 30 00:43:50 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N031820W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 18 08:49:09 2020  
Response via : Initial Calibration

