

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032720\
 Data File : VN060741.D
 Acq On : 27 Mar 2020 14:53
 Operator : JC/MD
 Sample : PB127814ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127814ZHE#06

Quant Time: Mar 30 00:50:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	203296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	341584	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133513	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
35) Dibromofluoromethane	7.57	113	100624	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
50) Toluene-d8	10.08	98	423391	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	159078	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
Target Compounds						
16) Acetone	3.78	43	5660	5.726	ug/l	96
20) Methylene Chloride	4.53	84	3339	1.373	ug/l	95
43) Isopropyl Acetate	8.14	43	63934	10.847	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032720\
Data File : VN060741.D
Acq On : 27 Mar 2020 14:53
Operator : JC/MD
Sample : PB127814ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127814ZHE#06

Quant Time: Mar 30 00:50:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

