

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032720\
 Data File : VN060736.D
 Acq On : 27 Mar 2020 13:00
 Operator : JC/MD
 Sample : PB127814ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127814ZHE#01

Quant Time: Mar 30 00:38:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	179633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	304435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	286278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	130874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	118000	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
35) Dibromofluoromethane	7.57	113	89867	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	10.08	98	372503	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.40	95	140592	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
Target Compounds						
16) Acetone	3.78	43	5561	6.366	ug/l	95
20) Methylene Chloride	4.52	84	3997	1.860	ug/l	96
43) Isopropyl Acetate	8.14	43	61043	11.620	ug/l #	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032720\
Data File : VN060736.D
Acq On : 27 Mar 2020 13:00
Operator : JC/MD
Sample : PB127814ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127814ZHE#01

Quant Time: Mar 30 00:38:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00

Instrument :

MSVOA_N

ClientSampleId :


PB127814ZHE#01

Tot Ion:168 Resp: 179633

Ion Ratio Lower Upper




168 100

99 55.9 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#16

Acetone

Concen: 6.366 ug/l

RT: 3.78 min Scan# 625

Delta R.T. 0.00 min

Lab File: VN060736.D


Acq: 27 Mar 2020 13:00

Tgt Ion: 43 Resp: 5561

Ion Ratio Lower Upper



43 100




58 34.3 25.1 37.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#20
Methylene Chloride
Concen: 1.860 ug/l
RT: 4.52 min Scan# 855
Delta R.T. 0.00 min
Lab File: VN060736.D
Acq: 27 Mar 2020 13:00


Instrument :
MSVOA_N
ClientSampleId :
PB127814ZHE#01


Tot Ion:	84	Resp:	3997
Ion	Ratio	Lower	Upper
84	100		
49	135.1	103.3	154.9
51	43.3	31.6	47.4
86	66.5	52.0	78.0



#33
1,2-Dichloroethane-d4
Concen: 47.852 ug/l
RT: 8.00 min Scan# 1937
Delta R.T. -0.00 min
Lab File: VN060736.D
Acq: 27 Mar 2020 13:00

Tgt Ion:	65	Resp:	118000
Ion	Ratio	Lower	Upper
65	100		
67	52.8	0.0	107.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00

Instrument :

MSVOA_N

ClientSampleId :

PB127814ZHE#01


Tot Ion:114 Resp: 304435

Ion Ratio Lower Upper

114 100

63 22.0 0.0 43.8




88 16.3 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 48.819 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00


Tgt Ion:113 Resp: 89867

Ion Ratio Lower Upper

113 100

111 103.6 82.9 124.3



192 20.3 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 11.620 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00

Instrument :

MSVOA_N

ClientSampleId :

PB127814ZHE#01


Tot Ion: 43 Resp: 61043

Ion Ratio Lower Upper

43 100

61 19.0 20.7 31.1#




87 11.7 9.8 14.8



Abundance Ion 43.00 (42.70 to 43.70): VN060736.D (-1887) (-)

Ion 61.00 (60.70 to 61.70): VN060736.D (-1887) (-)

Ion 87.00 (86.70 to 87.70): VN060736.D (-1887) (-)



#50

Toluene-d8

Concen: 49.541 ug/l

RT: 10.08 min Scan# 2583

Delta R.T. 0.00 min

Lab File: VN060736.D


Acq: 27 Mar 2020 13:00

Tgt Ion: 98 Resp: 372503

Ion Ratio Lower Upper



98 100


100 65.4 51.8 77.8



Abundance Ion 98.00 (97.70 to 98.70): VN060736.D (-2490) (-)

Ion 100.00 (99.70 to 100.70): VN060736.D (-2490) (-)





#62

4-Bromofluorobenzene

Concen: 51.020 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00

Instrument :

MSVOA_N

ClientSampleId :

PB127814ZHE#01


Tot Ion: 95 Resp: 140592

Ion Ratio Lower Upper

95 100

174 77.7 0.0 157.6

176 75.8 0.0 150.6






Abundance Ion 94.90 (94.60 to 95.60): VN060736.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN060736.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN060736.D (-3290) (-)

12.40



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2993

Delta R.T. -0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00


Tgt Ion: 117 Resp: 286278

Ion Ratio Lower Upper

117 100

82 58.0 46.7 70.1

119 32.2 25.5 38.3






Abundance Ion 116.90 (116.60 to 117.60): VN060736.D (-2979) (-)

Ion 82.00 (81.70 to 82.70): VN060736.D (-2979) (-)

Ion 118.90 (118.60 to 119.60): VN060736.D (-2979) (-)

11.40





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN060736.D

Acq: 27 Mar 2020 13:00

Instrument :

MSVOA_N

ClientSampleId :

PB127814ZHE#01


Tot Ion:152 Resp: 130874

Ion Ratio Lower Upper

152 100

115 59.6 30.2 90.6

150 155.3 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

