

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077200.D
 Acq On : 27 Mar 2023 13:28
 Operator : JC\MD
 Sample : VN0327WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0327WBL01

Quant Time: Mar 28 06:42:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 24 05:55:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.825	128	39913	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	252990	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	258568	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	120630	29.760	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.200%
60) 4-Bromofluorobenzene	12.848	95	112470	24.213	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.700%
63) Toluene-d8	10.572	98	281167	30.462	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.533%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077200.D
Acq On : 27 Mar 2023 13:28
Operator : JC\MD
Sample : VN0327WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0327WBL01

Quant Time: Mar 28 06:42:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 24 05:55:38 2023
Response via : Initial Calibration

