

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077206.D
 Acq On : 27 Mar 2023 17:07
 Operator : JC\MD
 Sample : VN0327WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0327WBL02

Quant Time: Mar 28 07:28:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	397649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	698592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	615006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	216372	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	300044	48.259	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.520%
35) Dibromofluoromethane	8.172	113	231800	48.900	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.800%
50) Toluene-d8	10.572	98	859351	48.381	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.854	95	246075	42.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077206.D
Acq On : 27 Mar 2023 17:07
Operator : JC\MD
Sample : VN0327WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0327WBL02

Quant Time: Mar 28 07:28:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

